


Operation of the Cobra Autosampler on the 8610V gas chromatograph

The Cobra Autosampler mounts on the top of the 8610V GC.



The Cobra keypad should be attached to the autosampler as shown.



There is a clamping bracket on the keypad. Loosen the two 5/32" hex screws holding the clamping bracket so there is a slight gap.

5/32" hex wrench

