


# Operation of the Cobra Autosampler on the 8610V gas chromatograph

Connect the keypad to the jack on the back of the Cobra.



Connect the remote control cable to the jack on the back of the Cobra.



Connect the other end of the remote control cable to the jack on the GC.

The remote control cable should be wired as shown.

