

Operation of the Cobra Autosampler on the 8610V gas chromatograph

Download the Cobra manual from the SRI website www.srigc.com

Use the keypad to configure the Cobra to your particular GC. You will have to tell the autosampler the position of the injector, sample bottles, waste and rinse bottles etc. The Cobra manual explains the meanings of the various parameters.

Before inserting the syringe, verify that the Cobra is aligned properly. Otherwise the syringe needle can be bent.

